
Midmark M11 - v1.0.0
Total Cycles: 1011

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2021 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 179.6 F 2.7
4:00 186.8 F 6.5
6:00 216.3 F 11.4
8:00 227.8 F 6.2
10:00 238.6 F 11.5
12:00 235.7 F 8.7
14:00 241.9 F 10.4
16:00 246.7 F 13.5
18:00 258.8 F 20.0
20:00 269.5 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.7
2:00 271.4 F 28.2
3:00 272.4 F 28.8
4:00 271.5 F 28.1

Min 271.3 F 28.1
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 227.0 F 0.0
10:00 211.8 F 0.0
15:00 210.4 F 0.0
20:00 212.5 F 0.0
25:00 225.2 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

07/14/2021 12:51 PM
ACCEPTED: X
REJECTED:
Operator: U01