
Midmark M11 - v1.0.0
Total Cycles: 1337

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2021 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 77.4 F 0.0
4:00 87.3 F 0.0
6:00 105.4 F 0.1
8:00 133.3 F 0.5
10:00 160.6 F 2.1
12:00 180.7 F 4.2
14:00 217.8 F 6.9
16:00 230.0 F 10.4
18:00 228.4 F 8.5
20:00 232.1 F 8.4
22:00 241.5 F 12.7
24:00 236.4 F 9.5
26:00 242.5 F 9.8
28:00 245.7 F 13.2
30:00 256.6 F 19.1
32:00 266.7 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.0
2:00 272.0 F 28.9
3:00 272.4 F 28.9
4:00 272.3 F 29.1

Min 271.5 F 28.5
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 203.6 F 0.0
10:00 205.4 F 0.0
15:00 207.0 F 0.0
20:00 201.4 F 0.0
25:00 200.6 F 0.0
30:00 198.9 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:35
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:11:49
CYCLE COMPLETE

11/08/2021 08:57 AM
ACCEPTED: X
REJECTED:
Operator: U01