
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001224
13:30 19/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:35
STERILIZING 4:50
136.0 C 220kPa 4:50
136.4 C 223kPa 5:20
135.8 C 220kPa 5:50
136.3 C 224kPa 6:20
136.0 C 222kPa 6:50
136.1 C 223kPa 7:20
136.4 C 225kPa 7:50
136.4 C 225kPa 8:20
136.4 C 225kPa 8:20
VENTING 8:20
AIR DRYING 9:13
CYCLE COMPLETE 22:24

Digital Signature #
86124047E1034D4C
