
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001553
14:29 11/03/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 5:30
STERILIZING 7:00
136.0 C 220kPa 7:00
136.1 C 218kPa 7:30
136.3 C 222kPa 8:00
136.3 C 222kPa 8:30
136.4 C 219kPa 9:00
136.5 C 220kPa 9:30
136.4 C 223kPa 10:00
136.4 C 223kPa 10:30
VENTING 10:30
AIR DRYING 11:33
CYCLE COMPLETE 33:08

Digital Signature #
53BB3EF6CD3DD40F
