
Midmark M11 - v1.0.5
Total Cycles: 820

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 11:21 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	115.8 F	0.0
2:00	118.6 F	0.8
4:00	129.3 F	3.5
6:00	166.7 F	8.3
8:00	199.3 F	6.6
10:00	223.1 F	8.6
12:00	234.0 F	9.0
14:00	244.5 F	13.0
16:00	245.6 F	13.2
18:00	253.1 F	16.4

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 247F
CHAMBER PRESS 13.3PSI
** ITEMS NOT STERILE **

02/22/2022 11:42 AM