
Midmark M11 - v1.0.5
Total Cycles: 4429

Name: RITTER2
Sterilizer ID: V2313228
03/14/2025 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.9 F 0.0
2:00 95.4 F 0.4
4:00 100.3 F 1.4
6:00 121.7 F 3.6
8:00 155.6 F 6.3
10:00 178.5 F 9.3
12:00 195.8 F 12.6
14:00 215.8 F 5.6
16:00 223.7 F 9.4
18:00 233.1 F 13.3
20:00 231.4 F 9.3
22:00 241.2 F 11.2
24:00 241.4 F 12.0
26:00 252.0 F 17.5
28:00 262.4 F 23.4
30:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 30.0
1:00 272.0 F 29.9
2:00 271.6 F 30.0
3:00 271.4 F 30.1
4:00 271.3 F 30.2

Min 271.3 F 29.9
Max 272.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 181.1 F 0.0
10:00 190.5 F 0.0
15:00 196.6 F 0.0
20:00 200.1 F 0.0
25:00 204.1 F 0.0
30:00 206.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

03/14/2025 12:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____