
Midmark M11 - v1.0.0
Total Cycles: 3411

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 143.8 F 0.0
4:00 153.1 F 1.1
6:00 181.4 F 4.5
8:00 198.9 F 8.7
10:00 216.1 F 11.5
12:00 227.3 F 7.8
14:00 238.2 F 12.7
16:00 237.3 F 9.8
18:00 237.1 F 8.8
20:00 248.4 F 14.2
22:00 259.9 F 20.6
24:00 270.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 272.1 F 28.7
3:00 271.7 F 28.2
4:00 272.5 F 28.9

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 207.7 F 0.0
10:00 229.1 F 0.0
15:00 225.2 F 0.0
20:00 221.6 F 0.0
25:00 220.5 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:48
CYCLE COMPLETE

06/20/2023 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____