
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005759
7:46 15/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:56
STERILIZING 4:00
136.5 C 222kPa 4:00
136.2 C 222kPa 4:30
136.2 C 219kPa 5:00
136.4 C 221kPa 5:30
136.1 C 222kPa 6:00
136.2 C 222kPa 6:30
136.4 C 223kPa 7:00
136.4 C 222kPa 7:30
136.4 C 222kPa 7:30
VENTING 7:30
AIR DRYING 8:21
CYCLE COMPLETE 8:27

Digital Signature #
4B78F0FCE43B5639
