

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5103

Name: RITTER 1  
Sterilizer ID: V2178198  
11/18/2024 01:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.9 F 0.0  
2:00 132.9 F 0.0  
4:00 146.1 F 0.2  
6:00 177.8 F 0.6  
8:00 199.6 F 2.9  
10:00 216.3 F 6.3  
12:00 223.6 F 10.2  
14:00 233.7 F 8.9  
16:00 236.6 F 10.4  
18:00 236.1 F 8.6  
20:00 245.7 F 13.2  
22:00 244.3 F 12.4  
24:00 255.9 F 18.5  
26:00 266.3 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.2 F 28.9  
2:00 272.3 F 28.8  
3:00 271.8 F 28.4  
4:00 271.3 F 28.1

Min 271.3 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.3 F 0.0  
5:00 200.1 F 0.0  
10:00 198.9 F 0.0  
15:00 215.8 F 0.0  
20:00 232.7 F 0.0  
25:00 228.6 F 0.0  
30:00 232.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 27:49  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:06:31  
CYCLE COMPLETE

11/18/2024 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_