
Midmark M11 - v1.0.5
Total Cycles: 1185

Name: RITTER2
Sterilizer ID: V2313228
06/13/2022 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 75.5 F 0.0
4:00 102.2 F 0.0
6:00 121.8 F 0.1
8:00 159.2 F 1.7
10:00 179.2 F 5.4
12:00 195.1 F 9.6
14:00 209.2 F 8.9
16:00 222.7 F 7.0
18:00 233.9 F 11.3
20:00 232.3 F 7.9
22:00 245.0 F 12.9
24:00 244.1 F 12.5
26:00 256.6 F 19.1
28:00 267.3 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.7 F 28.5
2:00 271.8 F 28.5
3:00 272.0 F 28.7
4:00 272.4 F 28.9

Min 271.6 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.3 F 0.0
5:00 228.6 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 29:19
STERILIZING: 4:00
VENTING: 2:42
DRYING: 9:08
TOTAL CYCLE: 00:46:11

06/13/2022 08:13 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____