

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1686

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2022 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.4 F 0.0  
2:00 167.3 F 0.1  
4:00 177.6 F 1.1  
6:00 203.4 F 5.8  
8:00 225.8 F 11.2  
10:00 228.7 F 7.5  
12:00 240.1 F 13.2  
14:00 240.7 F 11.5  
16:00 240.0 F 10.7  
18:00 254.1 F 17.8  
20:00 266.1 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.0 F 29.0  
2:00 272.6 F 29.6  
3:00 271.5 F 28.7  
4:00 272.4 F 29.4

Min 271.3 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 229.5 F 0.0  
10:00 226.8 F 0.0  
15:00 224.0 F 0.0  
20:00 223.1 F 0.0  
25:00 223.3 F 0.0  
30:00 221.0 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:32  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

11/07/2022 11:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_