
Midmark M11 - v1.0.5
Total Cycles: 4118

Name: RITTER2
Sterilizer ID: V2313228
11/14/2024 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.7 F 0.0
2:00 141.8 F 0.0
4:00 148.5 F 1.6
6:00 182.3 F 5.0
8:00 208.1 F 9.4
10:00 217.2 F 13.5
12:00 226.8 F 7.6
14:00 238.3 F 12.1
16:00 234.5 F 9.0
18:00 241.0 F 11.1
20:00 245.1 F 13.0
22:00 256.4 F 19.0
24:00 266.4 F 25.3
26:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.8 F 29.2
2:00 272.4 F 29.4
3:00 272.8 F 29.6
4:00 272.1 F 29.8

Min 271.4 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
5:00 244.0 F 0.0
10:00 223.0 F 0.0
15:00 182.8 F 0.0
20:00 194.5 F 0.0
25:00 199.2 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:24
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:02
CYCLE COMPLETE

11/14/2024 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____