
Midmark M11 - v1.0.0
Total Cycles: 3919

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2023 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.6 F 0.0
2:00 196.8 F 3.9
4:00 206.4 F 9.4
6:00 222.6 F 6.7
8:00 233.7 F 10.6
10:00 233.3 F 7.8
12:00 241.7 F 10.8
14:00 247.0 F 13.7
16:00 260.3 F 20.9
18:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 271.8 F 28.6
3:00 272.3 F 28.7
4:00 271.6 F 28.4

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 228.3 F 0.0
10:00 233.5 F 0.0
15:00 227.8 F 0.0
20:00 228.1 F 0.0
25:00 227.4 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

10/18/2023 12:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____