
Midmark M11 - v1.0.5
Total Cycles: 1825

Name: RITTER2
Sterilizer ID: V2313228
12/08/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 181.0 F 4.0
4:00 194.3 F 9.0
6:00 212.9 F 7.6
8:00 227.0 F 9.3
10:00 234.9 F 9.4
12:00 243.0 F 13.1
14:00 243.3 F 12.3
16:00 256.5 F 19.3
18:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.8
2:00 272.1 F 29.2
3:00 271.5 F 29.2
4:00 272.6 F 29.9

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 231.0 F 0.0
10:00 230.0 F 0.0
15:00 226.1 F 0.0
20:00 232.5 F 0.0
25:00 220.8 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

12/08/2022 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____