

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2360

Name: RITTER2  
Sterilizer ID: V2313228  
05/12/2023 10:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
2:00 187.0 F 3.1  
4:00 200.2 F 7.9  
6:00 218.1 F 13.6  
8:00 228.7 F 9.4  
10:00 234.2 F 9.2  
12:00 243.3 F 13.3  
14:00 243.4 F 12.5  
16:00 256.3 F 19.3  
18:00 267.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.9 F 29.8  
2:00 271.9 F 29.0  
3:00 271.9 F 29.6  
4:00 272.7 F 29.6  
  
Min 271.2 F 28.8  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 227.2 F 0.0  
10:00 235.7 F 0.0  
15:00 228.8 F 0.0  
20:00 224.8 F 0.0  
25:00 227.5 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 19:22  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

05/12/2023 11:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_