
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000814
16:13 24/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:42
PRESSURIZING 6:34
STERILIZING 8:10
136.0 C 220kPa 8:10
135.5 C 217kPa 8:40
136.2 C 222kPa 9:10
136.4 C 223kPa 9:40
135.9 C 220kPa 10:10
136.4 C 222kPa 10:40
135.7 C 219kPa 11:10
136.4 C 223kPa 11:40
136.4 C 223kPa 11:40
VENTING 11:40
AIR DRYING 12:46
CYCLE COMPLETE 42:46

Digital Signature #
5C7DFB551FCED79D
