
Midmark M11 - v1.0.0
Total Cycles: 5081

Name: RITTER 1
Sterilizer ID: V2178198
11/12/2024 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 82.6 F 0.0
2:00 84.6 F 0.0
4:00 107.1 F 0.1
6:00 140.8 F 0.7
8:00 169.5 F 0.8
10:00 191.4 F 1.0
12:00 204.8 F 3.5
14:00 212.5 F 6.2
16:00 221.6 F 9.3
18:00 231.5 F 13.1
20:00 232.4 F 8.4
22:00 243.1 F 13.1
24:00 240.6 F 10.7
26:00 236.6 F 8.5
28:00 250.1 F 14.8
30:00 260.8 F 20.8
32:00 270.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.8
1:00 272.6 F 28.4
2:00 271.5 F 27.8
3:00 272.7 F 28.5
4:00 271.7 F 27.8

Min 271.5 F 27.7
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 205.6 F 0.0
10:00 214.9 F 0.0
15:00 205.9 F 0.0
20:00 218.3 F 0.0
25:00 213.6 F 0.0
30:00 218.2 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:48
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:11:17
CYCLE COMPLETE

11/12/2024 03:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____