
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002170
13:39 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:04
136.4 C 224kPa 3:04
136.2 C 222kPa 3:34
136.4 C 225kPa 4:04
136.3 C 224kPa 4:34
136.4 C 223kPa 5:04
136.4 C 225kPa 5:34
136.1 C 222kPa 6:04
136.4 C 226kPa 6:34
136.4 C 226kPa 6:34
VENTING 6:34
AIR DRYING 7:23
CYCLE COMPLETE 37:24

Digital Signature #
205A2B3F053B58BF
