
Midmark M11 - v1.0.0
Total Cycles: 1046

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2021 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 181.4 F 3.9
4:00 197.3 F 8.4
6:00 215.4 F 9.9
8:00 229.5 F 8.4
10:00 241.8 F 11.9
12:00 241.5 F 11.5
14:00 239.4 F 10.1
16:00 253.4 F 16.9
18:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.8 F 28.8
2:00 271.6 F 28.2
3:00 272.9 F 28.9
4:00 272.0 F 28.3

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 205.3 F 0.0
10:00 210.4 F 0.0
15:00 227.6 F 0.0
20:00 215.6 F 0.0
25:00 222.6 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

07/22/2021 12:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____