
Midmark M11 - v1.0.5
Total Cycles: 4165

Name: RITTER2
Sterilizer ID: V2313228
11/28/2024 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.8 F 0.0
2:00 123.5 F 0.0
4:00 125.0 F 0.1
6:00 164.8 F 1.2
8:00 185.4 F 4.4
10:00 200.2 F 8.0
12:00 215.6 F 12.1
14:00 226.0 F 7.1
16:00 237.9 F 12.1
18:00 237.1 F 10.0
20:00 237.8 F 9.3
22:00 248.9 F 15.1
24:00 260.5 F 21.7
26:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.6
1:00 272.8 F 30.0
2:00 272.3 F 30.2
3:00 271.9 F 29.5
4:00 272.7 F 30.0

Min 271.8 F 29.4
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 224.9 F 0.0
10:00 221.9 F 0.0
15:00 207.6 F 0.0
20:00 196.6 F 0.0
25:00 194.7 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:15
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:55
CYCLE COMPLETE

11/28/2024 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____