
Midmark M11 - v1.0.0
Total Cycles: 4380

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 74.4 F 0.2
4:00 119.1 F 1.1
6:00 116.4 F 3.4
8:00 146.0 F 6.1
10:00 168.8 F 9.4
12:00 189.4 F 13.3
14:00 206.0 F 3.8
16:00 216.3 F 7.0
18:00 226.8 F 10.5
20:00 230.3 F 8.1
22:00 236.9 F 10.2
24:00 236.6 F 8.8
26:00 245.8 F 13.2
28:00 257.3 F 19.3
30:00 267.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.0
2:00 271.5 F 28.4
3:00 272.7 F 29.0
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 220.5 F 0.0
10:00 216.0 F 0.0
15:00 220.8 F 0.0
20:00 220.8 F 0.0
25:00 227.4 F 0.0
30:00 235.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

03/15/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____