
Midmark M11 - v1.0.5
Total Cycles: 1294

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 06:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 75.0 F 0.0
4:00 83.7 F 0.0
6:00 102.2 F 0.0
8:00 148.5 F 0.7
10:00 174.6 F 1.5
12:00 195.2 F 4.7
14:00 215.4 F 8.1
16:00 221.6 F 11.8
18:00 226.1 F 6.5
20:00 237.7 F 11.0
22:00 235.1 F 7.9
24:00 244.5 F 12.0
26:00 244.8 F 12.8
28:00 256.2 F 19.1
30:00 266.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.1
2:00 271.8 F 29.2
3:00 271.9 F 28.8
4:00 272.3 F 29.5

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 220.5 F 0.0
10:00 236.1 F 0.0
15:00 216.2 F 0.0
20:00 206.2 F 0.0
25:00 200.7 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:32
CYCLE COMPLETE

07/14/2022 07:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____