
Midmark M11 - v1.0.0
Total Cycles: 1126

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2021 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.8 F 0.0
2:00 155.5 F 1.3
4:00 168.3 F 3.9
6:00 183.4 F 7.6
8:00 218.4 F 12.5
10:00 220.0 F 6.4
12:00 230.1 F 11.2
14:00 232.1 F 7.6
16:00 242.8 F 13.5
18:00 243.6 F 11.7
20:00 255.7 F 18.0
22:00 266.5 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.1 F 29.0
2:00 271.2 F 28.3
3:00 271.9 F 29.2
4:00 272.0 F 28.7

Min 271.1 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 219.2 F 0.0
10:00 225.7 F 0.0
15:00 226.3 F 0.0
20:00 228.9 F 0.0
25:00 229.1 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:00
CYCLE COMPLETE

08/27/2021 02:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____