
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000149
10:01 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:02
STERILIZING 4:02
136.0 C 220kPa 4:02
136.3 C 219kPa 4:32
136.3 C 222kPa 5:02
136.4 C 223kPa 5:32
136.0 C 220kPa 6:02
136.4 C 225kPa 6:32
136.4 C 224kPa 7:02
136.4 C 221kPa 7:32
136.4 C 221kPa 7:32
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 8:34

Digital Signature #
F7807229D932D18A
