

-----  
Midmark M11 - v1.0.5  
Total Cycles: 397

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2021 12:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.4 F 0.0  
2:00 100.6 F 0.7  
4:00 123.2 F 2.8  
6:00 152.6 F 5.7  
8:00 176.5 F 9.1  
10:00 196.0 F 12.8  
12:00 212.8 F 6.2  
14:00 223.1 F 10.4  
16:00 228.5 F 8.4  
18:00 237.7 F 11.6  
20:00 236.6 F 9.7  
22:00 248.8 F 15.6  
24:00 259.7 F 21.6  
26:00 269.6 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 272.7 F 30.1  
2:00 271.6 F 29.4  
3:00 272.8 F 30.2  
4:00 271.9 F 29.6

Min 271.5 F 29.3  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
5:00 219.7 F 0.0  
10:00 201.6 F 0.0  
15:00 198.9 F 0.0  
20:00 199.1 F 0.0  
25:00 199.3 F 0.0  
30:00 194.5 F 0.0

Drying Complete

FILLING: 0:53  
HEATING: 27:00  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

09/27/2021 01:58 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01