
Midmark M11 - v1.0.5
Total Cycles: 3906

Name: RITTER2
Sterilizer ID: V2313228
09/04/2024 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.2 F 0.0
2:00 107.1 F 0.0
4:00 110.3 F 0.5
6:00 133.1 F 2.4
8:00 164.4 F 5.5
10:00 190.8 F 8.6
12:00 203.1 F 12.5
14:00 217.7 F 6.2
16:00 225.6 F 10.3
18:00 232.7 F 8.6
20:00 239.1 F 11.8
22:00 237.9 F 9.8
24:00 249.6 F 15.6
26:00 259.8 F 21.5
28:00 269.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.9 F 30.0
2:00 272.8 F 30.4
3:00 271.9 F 29.7
4:00 272.8 F 30.1

Min 271.8 F 29.7
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 220.2 F 0.0
10:00 206.3 F 0.0
15:00 210.9 F 0.0
20:00 229.9 F 0.0
25:00 231.2 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

09/04/2024 12:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____