

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001204  
14:14 03/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 3:55  
STERILIZING 5:03  
136.0 C 221kPa 5:03  
136.4 C 225kPa 5:33  
136.3 C 224kPa 6:03  
136.2 C 224kPa 6:33  
136.4 C 225kPa 7:03  
136.4 C 224kPa 7:33  
136.3 C 224kPa 8:03  
136.3 C 224kPa 8:33  
136.3 C 224kPa 8:33  
VENTING 8:33  
AIR DRYING 9:24  
CYCLE COMPLETE 24:24  
-----

Digital Signature #  
A469ACEA7C449C64  
-----