
Midmark M11 - v1.0.0
Total Cycles: 3449

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 07:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 73.9 F 0.0
4:00 83.1 F 0.0
6:00 114.1 F 0.4
8:00 143.7 F 1.8
10:00 169.8 F 3.7
12:00 196.3 F 6.3
14:00 202.2 F 10.1
16:00 218.5 F 8.2
18:00 222.7 F 7.3
20:00 233.0 F 11.3
22:00 229.7 F 7.2
24:00 243.6 F 12.9
26:00 240.7 F 11.0
28:00 253.3 F 17.1
30:00 263.9 F 23.4
32:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.2
2:00 271.7 F 28.3
3:00 272.2 F 29.2
4:00 272.3 F 28.8

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 229.7 F 0.0
10:00 219.1 F 0.0
15:00 224.7 F 0.0
20:00 224.6 F 0.0
25:00 228.8 F 0.0
30:00 225.3 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:08
CYCLE COMPLETE

06/27/2023 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____