

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2735

Name: RITTER2  
Sterilizer ID: V2313228  
08/04/2023 02:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.6 F 0.0  
2:00 85.7 F 0.0  
4:00 105.8 F 0.0  
6:00 133.6 F 0.1  
8:00 164.5 F 2.6  
10:00 185.0 F 5.5  
12:00 201.6 F 8.8  
14:00 214.6 F 12.3  
16:00 221.8 F 6.3  
18:00 231.8 F 10.4  
20:00 233.6 F 8.6  
22:00 242.7 F 12.5  
24:00 241.1 F 11.1  
26:00 252.7 F 17.1  
28:00 263.0 F 23.1  
30:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.8  
1:00 272.5 F 29.5  
2:00 271.4 F 28.7  
3:00 272.2 F 29.3  
4:00 271.3 F 28.9

Min 271.3 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 226.6 F 0.0  
10:00 192.0 F 0.0  
15:00 187.0 F 0.0  
20:00 184.9 F 0.0  
25:00 182.7 F 0.0  
30:00 181.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:13  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:08:39  
CYCLE COMPLETE

08/04/2023 03:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_