
Midmark M11 - v1.0.0
Total Cycles: 1019

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 181.4 F 3.6
4:00 210.4 F 8.0
6:00 218.1 F 13.5
8:00 228.3 F 8.1
10:00 242.7 F 13.4
12:00 240.6 F 11.1
14:00 238.5 F 9.6
16:00 252.4 F 16.4
18:00 264.0 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.4 F 28.3
3:00 272.7 F 28.9
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 229.4 F 0.0
10:00 229.1 F 0.0
15:00 227.1 F 0.0
20:00 227.3 F 0.0
25:00 232.2 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

07/15/2021 01:05 PM
ACCEPTED: X
REJECTED:
Operator: U02