
Midmark M11 - v1.0.0
Total Cycles: 5424

Name: RITTER 1
Sterilizer ID: V2178198
03/19/2025 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.4 F 0.0
2:00 102.3 F 0.3
4:00 106.6 F 1.9
6:00 133.7 F 4.5
8:00 165.1 F 7.6
10:00 185.7 F 10.9
12:00 209.5 F 6.0
14:00 216.4 F 6.1
16:00 223.6 F 9.9
18:00 232.0 F 8.7
20:00 237.1 F 10.4
22:00 236.5 F 8.5
24:00 246.5 F 13.8
26:00 257.4 F 19.6
28:00 267.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.5
2:00 272.0 F 28.9
3:00 271.5 F 29.1
4:00 272.4 F 29.4

Min 271.2 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 195.5 F 0.0
10:00 204.3 F 0.0
15:00 210.2 F 0.0
20:00 219.4 F 0.0
25:00 228.0 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:31
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:08:13
CYCLE COMPLETE

03/19/2025 04:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____