
Midmark M11 - v1.0.5
Total Cycles: 142

Name: RITTER2
Sterilizer ID: V2313228
06/25/2021 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 74.8 F 0.2
4:00 80.2 F 1.2
6:00 103.0 F 3.4
8:00 139.1 F 6.5
10:00 171.9 F 10.2
12:00 199.6 F 6.4
14:00 207.8 F 6.7
16:00 220.2 F 10.7
18:00 226.9 F 7.5
20:00 234.4 F 12.0
22:00 236.0 F 9.7
24:00 248.2 F 15.8
26:00 259.4 F 22.1
28:00 269.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.9
1:00 271.7 F 30.0
2:00 272.1 F 30.4
3:00 271.9 F 30.3
4:00 271.7 F 30.2

Min 271.2 F 29.8
Max 272.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 232.7 F 0.0
10:00 215.7 F 0.0
15:00 235.5 F 0.0
20:00 230.9 F 0.0
25:00 226.2 F 0.0
30:00 235.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:36
STERILIZING: 4:00
VENTING: 2:51
DRYING: 30:00
TOTAL CYCLE: 01:06:26
CYCLE COMPLETE

06/25/2021 10:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____