

-----  
Midmark M11 - v1.0.5  
Total Cycles: 712

Name: RITTER2  
Sterilizer ID: V2313228  
01/20/2022 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.0 F 0.0  
2:00 171.9 F 3.4  
4:00 194.5 F 8.4  
6:00 213.2 F 9.6  
8:00 225.1 F 9.4  
10:00 233.6 F 9.0  
12:00 242.8 F 13.4  
14:00 244.2 F 12.7  
16:00 257.4 F 19.8  
18:00 269.2 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 271.9 F 29.4  
2:00 272.4 F 29.4  
3:00 271.5 F 28.8  
4:00 272.0 F 29.6  
  
Min 271.3 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 212.5 F 0.0  
10:00 218.6 F 0.0  
15:00 205.3 F 0.0  
20:00 202.8 F 0.0  
25:00 211.6 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 18:58  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:39  
CYCLE COMPLETE

01/20/2022 11:01 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_