

-----  
Midmark M11 - v1.0.5  
Total Cycles: 267

Name: RITTER2  
Sterilizer ID: V2313228  
08/05/2021 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.8 F 0.0  
2:00 148.8 F 2.5  
4:00 170.2 F 6.6  
6:00 197.0 F 11.6  
8:00 216.4 F 5.4  
10:00 228.1 F 10.5  
12:00 232.1 F 8.0  
14:00 244.5 F 13.2  
16:00 243.0 F 12.0  
18:00 255.4 F 18.5  
20:00 266.5 F 25.3  
22:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.6 F 29.6  
2:00 272.1 F 29.7  
3:00 272.6 F 29.5  
4:00 272.0 F 29.7

Min 271.7 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.7 F 0.0  
5:00 230.1 F 0.0  
10:00 188.8 F 0.0  
15:00 184.2 F 0.0  
20:00 179.5 F 0.0  
25:00 176.0 F 0.0  
30:00 175.3 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 22:19  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

08/05/2021 01:11 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02