
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000580
12:03 05/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:29
STERILIZING 4:28
136.0 C 220kPa 4:28
136.3 C 224kPa 4:58
136.4 C 223kPa 5:28
136.3 C 224kPa 5:58
136.3 C 223kPa 6:28
136.4 C 224kPa 6:58
136.4 C 224kPa 7:28
136.2 C 221kPa 7:58
136.2 C 221kPa 7:58
VENTING 7:59
AIR DRYING 8:55
CYCLE COMPLETE 8:59

Digital Signature #
1B102F7753E4EC9A
