
Midmark M11 - v1.0.0
Total Cycles: 1221

Name: RITTER 1
Sterilizer ID: V2178198
09/30/2021 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.1 F 0.0
2:00 131.8 F 0.1
4:00 155.1 F 1.8
6:00 185.7 F 5.9
8:00 215.7 F 10.9
10:00 223.2 F 4.7
12:00 233.7 F 9.8
14:00 236.3 F 8.7
16:00 243.5 F 12.0
18:00 239.5 F 9.9
20:00 253.1 F 16.6
22:00 264.5 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.5 F 29.0
2:00 271.8 F 28.3
3:00 272.5 F 28.9
4:00 271.5 F 28.5

Min 270.9 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 240.3 F 0.0
10:00 228.8 F 0.0
15:00 194.2 F 0.0
20:00 181.2 F 0.0
25:00 191.9 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

09/30/2021 09:45 AM
ACCEPTED: X
REJECTED:
Operator: U02