
Midmark M11 - v1.0.0
Total Cycles: 1890

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 162.6 F 2.1
4:00 175.9 F 5.9
6:00 197.7 F 10.8
8:00 213.3 F 4.5
10:00 229.1 F 10.4
12:00 229.9 F 7.0
14:00 242.6 F 12.9
16:00 241.1 F 11.2
18:00 254.5 F 18.0
20:00 266.2 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.5 F 28.6
3:00 272.8 F 29.4
4:00 272.1 F 28.7

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 195.6 F 0.0
10:00 204.0 F 0.0
15:00 217.7 F 0.0
20:00 229.8 F 0.0
25:00 229.1 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

05/02/2022 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____