
Midmark M11 - v1.0.0
Total Cycles: 2577

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 165.3 F 3.8
4:00 182.4 F 8.3
6:00 202.3 F 13.2
8:00 219.9 F 6.7
10:00 231.3 F 11.6
12:00 232.3 F 8.3
14:00 240.6 F 11.0
16:00 245.2 F 13.0
18:00 257.9 F 19.8
20:00 269.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.2
2:00 271.6 F 28.4
3:00 272.4 F 29.4
4:00 272.1 F 28.9

Min 271.4 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 214.6 F 0.0
10:00 232.4 F 0.0
15:00 229.7 F 0.0
20:00 228.5 F 0.0
25:00 229.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

11/17/2022 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____