
Midmark M11 - v1.0.5
Total Cycles: 4331

Name: RITTER2
Sterilizer ID: V2313228
02/04/2025 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 151.6 F 1.7
4:00 160.7 F 5.1
6:00 187.1 F 10.1
8:00 210.6 F 3.8
10:00 232.7 F 10.6
12:00 232.4 F 8.2
14:00 239.5 F 10.0
16:00 248.3 F 14.6
18:00 261.0 F 21.7
20:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.9
2:00 272.9 F 29.5
3:00 271.8 F 28.9
4:00 272.4 F 29.0

Min 271.5 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.8 F 0.0
Dry Cycle Aborted

FILLING: 0:59
HEATING: 20:24
STERILIZING: 4:00
VENTING: 2:60
DRYING: 0:21
TOTAL CYCLE: 00:28:45

02/04/2025 10:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____