
Midmark M11 - v1.0.0
Total Cycles: 1223

Name: RITTER 1
Sterilizer ID: V2178198
09/30/2021 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 184.3 F 3.5
4:00 208.4 F 7.8
6:00 220.2 F 11.7
8:00 230.9 F 8.5
10:00 238.7 F 11.2
12:00 242.2 F 11.6
14:00 240.0 F 10.1
16:00 254.0 F 17.2
18:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.4 F 28.1
3:00 272.7 F 28.8
4:00 271.7 F 28.1

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 224.3 F 0.0
10:00 214.6 F 0.0
15:00 204.0 F 0.0
20:00 224.6 F 0.0
25:00 221.1 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

09/30/2021 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____