
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000535
8:30 25/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:36
STERILIZING 4:59
136.0 C 227kPa 4:59
136.3 C 232kPa 5:29
136.2 C 231kPa 5:59
136.4 C 232kPa 6:29
136.4 C 230kPa 6:59
136.3 C 231kPa 7:29
136.4 C 231kPa 7:59
136.2 C 231kPa 8:29
136.2 C 231kPa 8:29
VENTING 8:29
AIR DRYING 9:27
CYCLE COMPLETE 24:27

Digital Signature #
6BD2587EE5D90243
