
Midmark M11 - v1.0.5
Total Cycles: 4397

Name: RITTER2
Sterilizer ID: V2313228
03/04/2025 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.2 F 0.0
2:00 86.7 F 0.1
4:00 96.6 F 1.5
6:00 129.4 F 4.1
8:00 158.6 F 7.3
10:00 177.1 F 10.5
12:00 215.3 F 6.0
14:00 222.7 F 6.5
16:00 228.7 F 10.1
18:00 235.5 F 9.7
20:00 237.6 F 10.4
22:00 236.1 F 8.7
24:00 246.1 F 13.7
26:00 256.5 F 19.4
28:00 266.4 F 25.5
30:00 272.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 271.7 F 29.7
2:00 271.8 F 29.8
3:00 271.8 F 29.7
4:00 271.7 F 29.6

Min 271.6 F 29.5
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 224.7 F 0.0
10:00 213.7 F 0.0
15:00 196.9 F 0.0
20:00 224.4 F 0.0
25:00 217.3 F 0.0
30:00 187.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

03/04/2025 01:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____