

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1858

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2022 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
2:00 200.5 F 4.6  
4:00 206.5 F 10.4  
6:00 225.5 F 6.5  
8:00 240.3 F 13.3  
10:00 241.8 F 11.8  
12:00 241.3 F 11.3  
14:00 255.9 F 18.9  
16:00 268.4 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.3 F 29.4  
2:00 271.3 F 28.6  
3:00 272.1 F 29.6  
4:00 272.0 F 29.2  
  
Min 271.0 F 28.6  
Max 272.4 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 232.1 F 0.0  
10:00 232.8 F 0.0  
15:00 227.5 F 0.0  
20:00 228.6 F 0.0  
25:00 228.8 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 3:04  
HEATING: 17:09  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:21  
CYCLE COMPLETE

12/19/2022 04:42 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_