
Midmark M11 - v1.0.5
Total Cycles: 2889

Name: RITTER2
Sterilizer ID: V2313228
09/15/2023 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 153.7 F 0.2
4:00 172.8 F 3.2
6:00 193.8 F 7.2
8:00 209.4 F 11.5
10:00 219.6 F 5.4
12:00 230.6 F 10.0
14:00 233.9 F 9.2
16:00 241.8 F 12.2
18:00 240.6 F 10.7
20:00 252.9 F 16.9
22:00 263.5 F 23.2
24:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.3 F 28.4
3:00 272.4 F 29.3
4:00 271.6 F 28.6

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 194.8 F 0.0
10:00 194.7 F 0.0
15:00 191.7 F 0.0
20:00 187.9 F 0.0
25:00 185.2 F 0.0
30:00 182.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

09/15/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____