
Midmark M11 - v1.0.5
Total Cycles: 4226

Name: RITTER2
Sterilizer ID: V2313228
12/18/2024 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.2 F 0.0
2:00 142.1 F 1.8
4:00 149.6 F 4.6
6:00 174.5 F 8.3
8:00 196.7 F 12.3
10:00 211.1 F 5.0
12:00 223.9 F 8.9
14:00 235.3 F 13.3
16:00 235.8 F 10.2
18:00 237.3 F 9.3
20:00 247.0 F 14.3
22:00 257.9 F 20.3
24:00 267.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 271.8 F 29.5
2:00 272.9 F 30.0
3:00 272.3 F 29.5
4:00 271.6 F 29.1

Min 271.5 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 195.8 F 0.0
10:00 197.6 F 0.0
15:00 196.4 F 0.0
20:00 190.7 F 0.0
25:00 185.6 F 0.0
30:00 191.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:19
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

12/18/2024 02:32 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____