

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3717

Name: RITTER 1  
Sterilizer ID: V2178198  
09/05/2023 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
2:00 181.3 F 1.9  
4:00 195.4 F 6.4  
6:00 213.0 F 10.9  
8:00 227.2 F 5.8  
10:00 237.9 F 11.4  
12:00 235.0 F 9.0  
14:00 240.7 F 9.9  
16:00 247.8 F 14.2  
18:00 259.6 F 20.6  
20:00 270.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 29.0  
2:00 271.2 F 28.2  
3:00 272.7 F 29.2  
4:00 272.1 F 28.6  
  
Min 271.1 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 229.5 F 0.0  
10:00 238.1 F 0.0  
15:00 228.8 F 0.0  
20:00 225.0 F 0.0  
25:00 224.1 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

09/05/2023 12:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_