
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000257
14:52 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:31
STERILIZING 4:42
136.4 C 225kPa 4:42
136.4 C 224kPa 5:12
136.4 C 224kPa 5:42
136.4 C 223kPa 6:12
136.3 C 224kPa 6:42
136.4 C 223kPa 7:12
136.2 C 224kPa 7:42
136.2 C 222kPa 8:12
136.2 C 222kPa 8:12
VENTING 8:12
AIR DRYING 9:04
CYCLE COMPLETE 9:25

Digital Signature #
C0AA2615861D093E
