
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001121
9:57 18/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:37
STERILIZING 4:42
136.5 C 224kPa 4:42
136.3 C 224kPa 5:12
136.2 C 220kPa 5:42
136.3 C 223kPa 6:12
136.4 C 225kPa 6:42
136.4 C 222kPa 7:12
136.4 C 225kPa 7:42
136.4 C 221kPa 8:12
136.4 C 221kPa 8:12
VENTING 8:12
AIR DRYING 9:13
CYCLE COMPLETE 9:14

Digital Signature #
B7E3A127F269FE14
