
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001109
14:26 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:28
STERILIZING 3:53
136.4 C 225kPa 3:53
136.4 C 225kPa 4:23
136.4 C 225kPa 4:53
136.2 C 224kPa 5:23
136.3 C 225kPa 5:53
136.4 C 222kPa 6:23
136.1 C 221kPa 6:53
136.1 C 221kPa 7:23
136.1 C 221kPa 7:23
VENTING 7:23
AIR DRYING 8:12
CYCLE COMPLETE 23:12

Digital Signature #
25B0557D8CE90F65
