
Midmark M11 - v1.0.0
Total Cycles: 5297

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2025 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.6 F 0.0
4:00 87.8 F 0.1
6:00 122.1 F 2.4
8:00 149.9 F 5.0
10:00 176.0 F 8.2
12:00 200.8 F 12.0
14:00 211.3 F 3.3
16:00 224.0 F 7.1
18:00 233.3 F 11.0
20:00 229.0 F 6.1
22:00 241.5 F 11.5
24:00 237.2 F 8.9
26:00 250.5 F 15.2
28:00 261.6 F 21.5
30:00 271.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.9
2:00 271.8 F 28.1
3:00 272.1 F 28.5
4:00 272.5 F 28.7

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 180.0 F 0.0
10:00 203.4 F 0.0
15:00 231.8 F 0.0
20:00 214.0 F 0.0
25:00 200.7 F 0.0
30:00 197.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:38
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:08:59
CYCLE COMPLETE

01/30/2025 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____